

# THYROGLOBULIN CONTROL – TRI LEVEL

Thyroglobulin (Tg)  
PRODUCT CODE: TG-300

LOT# TG(A-C)1G05

EXP: 07-08

Revision: 1

Date: 04-12-06

## INTENDED USE

The Thyroglobulin Controls are intended for use as an assayed quality control material to monitor the consistency of performance of laboratory test procedures associated with determination and monitoring of thyroglobulin levels. This product is human serum based, liquid, stabilized with preservatives and can be used with all RIA, EIA, ELISA or FIA methods.

## SUMMARY AND EXPLANATION

The use of quality control material to assist in the assessment of precision in the clinical laboratory is an integral part of laboratory practices. Controls that contain varied levels of analytes are necessary to insure precision and accuracy in immunoassay systems.

## REAGENTS

Monobind Inc.'s thyroglobulin controls are intended to be used in the exact manner as patient samples. The control is packaged as 6 vials of 3.0 ml (2 of each level). The analyte activities are adjusted to concentrations in the low, middle and high range in order to monitor the efficacy of the procedure in use.

## INSTRUCTIONS FOR USE

- 1). Bring the vials to room temperature before use.
- 2). Carefully unscrew and remove cap.
- 3). Store the unused portions at 2-8°C after each use.

## STORAGE, STABILITY AND DISPOSAL

The expiration date (<-20°C storage) is printed on each vial as well as on the outside container. After opening, any unused material is stable for 30 days when stored at 2-8°C (see below). Outdated material should be discarded as biohazardous component.

STORAGE	STABILITY	TEMPERATURE
Unopened	Printed on vial	2 – 8°C
Opened	Thirty (30) days	2 – 8°C

## EXPECTED VALUES

The expected ranges are given below;

### Low Control (A) Mean (ng/ml) Expected Range

Access™	1.6	0.5 – 2.7
Monobind ELISA	1.7	0.5 – 2.7
Monobind CLIA	1.8	0.5 – 2.7

### Normal Control (B) Mean (ng/ml) Expected Range

Access™	13	11 - 15
Monobind ELISA	14.6	11 - 18
Monobind CLIA	13.3	10 - 17

### High Control (C) Mean (ng/ml) Expected Range

Access™	60	50 - 70
Monobind ELISA	62.2	50 - 80
Monobind CLIA	60	45 - 75

Access is a trademark of Beckman-Coulter, Inc.

Each laboratory should establish their own ranges for the product used. A trend log should be maintained for batch to batch consistency of the test. Large variations from the established mean may result from (a) improper technique, (b) improper dilutions from the stated manufacturer's procedure or (c) modifications of the test system by the manufacturer.

## WARNING AND PRECAUTIONS

All products that contain human serum have been found to be non reactive for HIV 1&2, HIV-Ag, HBsAg, HCV and RPR by FDA required tests. Since no known test can offer complete assurance that infectious agents are absent, all human serum products should be handled as potentially hazardous and capable of transmitting disease. Good laboratory procedures for handling blood products can be found in the Center for Disease Control / National Institute of Health, "Biosafety in Microbiological and Biomedical Laboratories," 2nd Edition, 1988, HHS Publication No. (CDC) 88-8395.

## FOR IN VITRO DIAGNOSTIC USE

For Orders and Inquiries, please contact



Tel: 949-951-2665  
Fax: 949-951-3539  
Email: [info@monobind.com](mailto:info@monobind.com)  
On the Web: [www.monobind.com](http://www.monobind.com)



EC REP CEpartner4U, 3951 DB; 13.NL  
Tel: +31 (0) 6-516.536.26